

# Declaration of Conformity



Manufacturers Name: Gordon Ellis & Co.

Manufacturers Address: Trent Lane  
Castle Donington  
Derby  
DE74 2AT  
United Kingdom

GMDN Code: 31059

Authorised Representative Name: UNITED KINGDOM  
Gordon Ellis & Co. EUROPE  
Advena Ltd

Authorised Representative Address: Trent Lane  
Castle Donington  
Derby  
DE74 2AT  
United Kingdom Tower Business Centre, 2nd Floor  
Tower Street  
Swatar  
BKR 4013  
Malta

SRN (Single Registration Number): GB-MF-000012829 MT-AR-000000234

Basic UDI-DI: 5016181ASHBYLIFT73

Name of the Device(s): Ashby Lift Oval 3IN Raise "Ashby Lift Oval"  
Ashby Lift Straight 3IN Raise "Ashby Lift Straight"

Product Code(s): 66294  
66394

Intended Purpose : The device is intended to raise the seat height of a toilet to assist those with limited mobility as it reduces the movement required when sitting or standing and is used as an aid to daily living.

Classification: Class I

Conformity Assessment Route: Gordon Ellis & Co. uses the requirements for the CE-labelling of their products according the Regulations MDR 2017/745 - Article 19 Annex II and Annex III.

Gordon Ellis & Co. test their products in accordance with BS EN 21856:2022 - Assistive Products for Persons with Disability - General Requirements and Test Methods and Risk Assess to BS EN ISO 14971:2019 - Application of Risk Management to Medical Devices.

This declaration of conformity is issued under the sole responsibility of Gordon Ellis & Co. We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

All supporting documentation is retained at the premises of the manufacturer.

Signature:

A handwritten signature in black ink, appearing to be 'Ashley Kidd', written over a horizontal line.

Place and Date (DD/MM/YYYY) of Issue:

Location: Gordon Ellis & Co. Date: 30/01/2024

Name:

Ashley Kidd

Function:

QHSE Manager

Signing of this declaration of conformity authorises the manufacturer to affix the CE Mark to the product in accordance with the above directive.

